

METHOD NOTES



(sludge, soil, sediment, oil)

Sample Category: Environmental & Regulatory

Sample Requirements

Sample Weight: 0.5 g

Reagents: 10 mL Conc. HNO₃

Vessels: 35 mL glass, quartz, and Teflon™ lined

Microwave Heating Program

Control Style: Ramp To Temperature


Pressure Mode: Organic


Recommended Equipment




Discover SP-D

Stage	Temp (°C)	Ramp Time	Hold Time	Pressure (psi)	Power (W)	Stirring
1	165	2:00	0:00	350	300	Med
2	175	3:00	5:00	400	300	Med







Notes:

1. If using a vial other than the recommended choice, adjust sample size and pressure limit to values appropriate for the vessel chosen.

General Precautions:

This procedure is a reference for sample digestion using a CEM System and may need to be modified or changed to obtain the required results on your sample

www.cem.com • info@cem.com • 704-821-7015 • 800-726-3331