



Procedure

Weigh 0.5 g of sample into the digestion vessel. Add 10 mL HCl. Gently swirl the vessel to thoroughly mix the sample and acid.

Notes

This application can only be run in the iPrep vessel.

The maximum number of vessels that can be run for this method is 6.

Ensure that particles are well suspended in acid mixture prior to sealing vessel.

Stirring may be helpful to keep particles suspended in solution during digestion.

Recommended Equipment

MARS 6 iWave

Recommended Vessels

iPrep

Reagents

HCl

Max Sample Weight

0.5 g

Sample Type

HCl

Control Type

Ramp to Temperature

Method Type

Classic

Heating Program

Stage	Temp (°C)	*Ramp (mm:ss)	Hold (mm:ss)	Pressure (psi)	* Power (W)	Stirring
1	240	30:00	59:59	N/A	900-1800	N/A

* Ramp times and power may vary depending on the type and number of vessels.

Results

Sample was clear, colorless, and particle free upon dilution to 50 mL.

General Precaution

- This procedure is a reference point for sample digestion using a CEM system and may need to be modified or changed to obtain the required results on your sample.
- If using a vessel other than the recommended choice, adjust sample size and pressure limit to values appropriate for the vessel chosen.
- Manual venting of CEM vessels should be performed when wearing hand/eye/body protection and when the vessel contents are at or below room temperature to avoid the potential for chemical burns. Always point the vent hole away from the operator.
- If programming as One Touch, the ramp time and power will be automatically determined based on the number and type of vessels detected.