

Mars 6TM Method Note

Microwave Digestion of Urine - Human

Procedure

Transfer 4 mL of the sample into the digestion vessel. Add 4 mL of HNO_3 and 2 mL of H_2O_2 . Gently swirl the mixture and wait approximately 15 minutes before closing the vessel.

Notes

Take caution when mixing the acids and the sample. The sample may foam slightly.

 Recommended Equipment
 Recommended Vessels
 Reagents

 MARS 6
 EasyPrep Plus H2O2

 MARS 6 iWave
 EasyPrep Plus H2O2

 MARSXpress MARSXpress Plus
 MARSXpress Plus

Max Sample Weight	Sample Type	Control Type	Method Type
4 mL	Organic	Ramp to Temperature	One Touch

Heating Progra	am					
Stage	Temp (°C)	*Ramp (mm:ss)	Hold (mm:ss)	Pressure (psi)	* Power (W)	Stirring
1	200	15:00	15:00	800	900-1050	Off

^{*} Ramp times and power may vary depending on the type and number of vessels.

Results

Sample was clear, colorless, and particle free upon dilution to 50 mL.

General Precaution

- a) This procedure is a reference point for sample digestion using a CEM system and may need to be modified or changed to obtain the required results on your sample.
- b) If using a vessel other than the recommended choice, adjust sample size and pressure limit to values appropriate for the vessel chosen.
- c) The control / reference vessel must contain the largest and most reactive sample.
- d) Manual venting of CEM vessels should be performed when wearing hand/eye/body protection and when the vessel contents are at or below room temperature to avoid the potential for chemical burns. Always point the vent hole away from the operator.
- e) If programming as One Touch, the ramp time and power will be automatically determined based on the number and type of vessels detected.