

**Systems**ORACLE
SMART 6**Equipment**CEM Square Glass Fiber Pads
CEM Trac Film
ORACLE Station
Sample Press
Spatula**SMART 6 Settings**

Calculation Mode	Option	Display Mode
Moisture/Fat	Standard	Moisture

SMART 6 Categories

Ramp Time (sec)	Delta Weight (mg)	Delta Time (sec)	
30	5	20	
Am. Fan Speed (%)	Run Fan Speed (%)	Run Fan Off (sec)	Weight Comp. (On /Off)
0	50	30	On

ORACLE Parameters

Cooling Enabled	Temperature (°C)	QuickPrep™ Temp (°C)	QuickPrep™ Time (sec)	Warm Delay (sec)	Max Cool Time
Yes	60	70	30	5	1:30

Parameters

Stage	Mag (%)	IR (%)	Time (mm:ss)	Max Temp (°C)	Mode
1	100	0	3:30	160	Set Time

Sample Preparation

Blend sample using an industrial or consumer grade food processor until homogeneous.

Procedure

Tare two glass fiber sample pads. Apply 1.9-2.1 grams of sample on two pads, press with sample press. Place the pads with sample centered on balance pan. Close cover, and press "START". When prompted by the SMART 6, prepare the samples using the ORACLE Station and follow prompts on the display.

Spread Technique