



### Procedure

Weigh 0.1 g (0.05 mL if using a 4 mL vessel) of the sample into the digestion vessel. Add 1 mL of HNO<sub>3</sub>. Gently swirl the mixture and wait approximately 15 minutes before capping the vessel.

### Notes

#### Recommended Equipment

Discover SP-D Clinical

#### Recommended Vessels

10 mL Pyrex  
4 mL Pyrex

#### Reagents

HNO<sub>3</sub>

#### Max Sample Weight

0.1 g

#### Sample Type / Vent Program

Organic

#### Control Type

Ramp to Temperature

### Heating Program

Stage	Temp (°C)	*Ramp (mm:ss)	Hold (mm:ss)	Pressure (psi)	* Power (W)	Stirring
1	200	4:00	3:00	400	300	Med

\* Ramp times and power may vary depending on the type and number of vessels.

### Results

Sample was clear, colorless, and particle free upon dilution to 50 mL.

### General Precaution

- This procedure is a reference point for sample digestion using a CEM system and may need to be modified or changed to obtain the required results on your sample.
- If using a vessel other than the recommended choice, adjust sample size and pressure limit to values appropriate for the vessel chosen.